

Office of the Director,
INSTITUTE OF NEPHRO-UROLOGY
(Government of Karnataka - Autonomous Institute)
Victoria Hospital Campus,
Bangalore - 560 002. Red Buckets with Lid
Telefax : (080) 26706777
Email-nephrourology.institute@gmail.com

No: INU/ADMN/PROC/523/2022

Dated: 09/01/23

INVITATION OF QUOTATIONS FOR LAB CENTRIFUGE - CATA-II

1. M/s. Vijay Enterprises
#581, 60 Feet Road, 7th Block,
9th Phase, J.P Nagar,
Anjanapura Post,
Bangalore - 560 108.
2. M/s. Aims
#131 Surveyerstreet
Basavanagudi
Bangalore - 04
3. M/s. Hari Hara Instruments
Supplies
#69, 1st Floor, 2nd Main
Road, Palace Guttahalli,
Bangalore - 560 003
4. M/s. JLD Scientific Co.
#581, 60 Feet Road, 7th Block,
9th Phase, Anjanapura Post,
J.P Nagar,
Bangalore - 560 062
5. M/s. K.K Surgicals
No 40 1st Floor,
14th main 4th cross,
Maruthi Circle
Hanumath Nagar
Bangalore
6. M/s. Array Med,
404, 1st Floor, 1st Cross,
6th Main, HAL 2nd Stage,
Bangalore- 560 008.
7. M/s.BPL Limited,
11th KM, Arekere,
Bhanneghatta Road,
Bangalore - 560 076.
8. M/s. Imago System,
#27, "Sneha", 2nd Floor,
1st Cross, 1st Main,
Arakere Samrat Layout,
Off Bannerghatta Road,
Bangalore - 560 076.
9. M/s. Ayuh Meditech solution
Pvt. Ltd.
#8,Sriyal,2nd Floor, 8th 'A'
Cross,C.I.L Layout,
Cholanayakkana Halli,
R.T Nagar post,
Bangalore - 560 032.
- 10 Notice Board
INU
Bangalore.
- 11 Website
www.nephrourology.in
(Under Quotation Column)

Dear Sirs,

Sub: Invitation of quotation for supply of Lab Centifuge - reg.

1. Sealed competitive quotations are invited by the undersigned for the following items.

Sl No	Name of the Item	Specification	Qty
1.	Lab Centrifuge (Bench top centrifuge)	TECHNICAL: Stainless Steel centrifuge Chamber,easy to clean. Brushless Induction motor with variable frequency drive. Microprocessor controller with digital display. Stable speed output even under unstable voltage condition.	02 Nos.

		<p>Smooth & Soft start.</p> <p>Low sample temperature rise.</p> <p>Inverter fault detection without shutdown.</p> <p>7 Segment LED display of speed.</p> <p>Digital countdown timer & continuous run.</p> <p>Safety lid interlock to prevent lid opening during centrifugation.</p> <p>Imbalance detection & centrifugation stop with display of error.</p> <p>Dynamic brake for quick deceleration.</p> <p>Motor overload protection.</p> <p>Gas hinge to prevent door falling.</p> <p>Emergency lid lock release.</p> <p>Last set parameters recall (Useful for repetitive analysis).</p> <p>Wide variety of rotors & reduction adaptors .</p>	
--	--	--	--

Max. Speed	Rpm	6000
Max. RCF	'g'	5070
Max. Capacity	ml	400
Digital Timer Range	Min	0-59
W x D x H	Mm	380 x 470 x 300
Supply	220-240 Volts	50 Hz Single Phase

CENTRIFUGE HIGH SPEED WITH TACHOMETER---02 nos

GENERAL:

1. Should have user friendly microprocessor control with Interactive LCD display and soft touch keypads for fast and accurate setting of run parameters.
2. Should do timely calibration and maintenance with providing proper documentary evidence in accordance with NABL protocols.

Quoted Price:

- The tenderer/Vendor shall quote for items in the format of quotation attached;
 - a. The specifications should be drafted to permit the widest possible competition and at the same time present a clear statement of the required standards of workmanship, materials, and performances of the goods to be procured. Minimum functional specifications should be specified for equipment. Only if this is done will the objectives of economy, efficiency and fairness in procurement are realized, responsiveness of quotations be ensured and the subsequent task of evaluation of quotations facilitated.
 - b. Wherever the Goods are covered by Bureau of Indian Standards, the reference to the standards should be given. If ISI marking is available it should be specified.
 - c. Reference to Brand name and catalogue number should be avoided as far as possible. Where unavoidable, they should be followed by the words "or at least equivalent"
- All duties, taxes and other levies payable by the tenderer/Vendor (including Sales Tax on the finished goods) shall be included in the item rate.
- The rates quoted for each item shall fixed for the duration of the contract and shall not be subject to any adjustment.
- Rates for supply of partial quantity of an item are not acceptable.
- Corrections if any shall be made by crossing out, initialing, dating and rewriting.
- Cable of facsimile quotations are not acceptable.

2. Each tenderer/Vendor must submit only one quotation.

3. Validity of Quotations:

The Quotation shall remain valid for a period not less than 30 days after the deadline fixed for submission of quotations.

4. Evaluation of quotations :

The samples / Catalogue with technical specification should be accompanied along with quotations. The same will be tested and if the samples / Catalogue are found suitable only such quotations will be treated as responsive quotation.

- a. The Purchaser will evaluate and compare the quotations determined to be substantially responsive i.e., which are properly signed, and conform to the terms and conditions and specifications in the following manner:
- b. The evaluation will be done including the Sales tax. If the tenderer/Vendor has not included the Sales Tax in his quotation for the item rate, and has also not indicated the rate of Sales Tax applicable, the quoted rate will treated as though it is inclusive of the Sales tax and no extra payment for Sales Tax will be made;
- c. The evaluation would be done for all the items put together. The items for which no rates have been quoted would be treated as zero and the total amount would be computed accordingly. The tenderer/Vendor who has quoted for partial quantity of

any one or more item(s) would be treated as non-responsive. Purchaser will award the contract to the responsive tenderer, whose total cost for all the items put together is the lowest.

5. Award of Contract:

- a. The Purchaser will award the contract to the tenderer/Vendor whose quotation has been determined to be substantially responsive and who has offered the lowest price as per para 5(c) above only after testing the samples and subject to approval by the experts.
- b. The Purchaser reserves the right at the time of contract award to increase or decrease the quantities of any or all the items indicated in para 1 above by 25% without any change in the unit price or any other terms and conditions.
- c. The Purchaser prior to the expiration of the quotation validity period will notify the tenderer/Vendor whose quotation is accepted of the award of contract. The terms of the accepted offer shall be incorporated in the Purchase Order.
- d. Normal commercial warranty / guarantee shall be applicable to the supplied goods.
- e. Payment shall be made immediately after the delivery of the goods and their acceptance.
- f. Notwithstanding the above, the Purchaser reserves the right to accept or reject any quotations and to cancel the quotation process and reject all quotations at any time prior to the award of the contract.

6. Last date and time of receipt of quotations:

You are requested to submit the sealed quotations (with catalogue if any) superscribed on the envelope as "Quotations for Supply of Lab Centrifuge" Thursday Due on _____ " latest by 4PM hours on 12/01/23 (date).

We look forward to receiving your quotations.

Yours Sincerely,


Director,
Institute of Nephro-Urology,
Bangalore

Note:

1. If quotation is submitted by hand, acknowledgement should be obtained from the concerned authorized person in the institute.
 2. If the quotation is posted, it should be only by registered post with acknowledgement due.
 3. Quotations received after the due date will not be accepted.
- * Any violation of the above procedures the quotation even if it is received in the office shall not be considered.

FORMAT OF QUOTATION

Sl No	Description of Goods / equipment	Brief Specifications	Make / Manufacturer Name	Qty	Unit	Unit Rate (Rs) in figures	Unit Rate (Rs) in words	Total amount (in figures)
1.								
2.								
3.								
4.								
5.								
6.								
7.								
8.								
9.								
10.								

Gross total Cost : Rs _____ (in figures)

Rs _____ (in words)

- We agree to supply the above goods in accordance with the technical specifications for a total contract price of Rs _____ (in figures) (Rs _____ (in words), within the period specified in the Invitation for Quotations.
- We also confirm that the normal commercial warranty/guarantee of _____ months shall apply to the offered goods.

(Tenderer)

Name.....

Signature

Date

